


How To Do the Rapid Test for Malaria

Modified for training in the use of the **Generic Pf Test** for falciparum malaria




Collect:

- NEW unopened** test packet
- NEW unopened** alcohol swab
- NEW unopened** lancet
- NEW** pair of disposable gloves
- Buffer
- Timer



READ THESE INSTRUCTIONS CAREFULLY BEFORE YOU BEGIN.

- 1.** Check the expiry date on the test packet.



Expiry date

- 2.** Put on the gloves. Use new gloves for each patient



- 3.** Open the packet and remove:




- Test
- Capillary tube
- Desiccant sachet

- 4.** Write the patient's name on the test.



- 5.** Open the alcohol swab. Grasp the 4th finger on the patient's left hand. Clean the finger with the alcohol swab. Allow the finger to dry before pricking.




- 6.** Open the lancet. Prick patient's finger to get a drop of blood.




- 7.** Discard the lancet in the Sharps Box immediately after pricking finger. **Do not set the lancet down before discarding it.**



- 8.** Use the capillary tube to collect the drop of blood.




- 9.** Use the capillary tube to put the drop of blood into the square hole marked "A."



- 10.** Discard the capillary tube in the Sharps Box.



- 11.** Add buffer into the round hole marked "B."



Count correct number of drops

- 12.** Wait **15 minutes** after adding buffer.




- 13.** Read test results. **(NOTE: Do Not** read the test sooner than **15 minutes** after adding the buffer. You may get **FALSE** results.)

- 14.** How to read the test results:


POSITIVE

A line near letter "C" and a line near letter "T" means the patient is **POSITIVE** for malaria.



P. falciparum


The test is **positive** even if the line near "T" is **faint**.)



P. falciparum (faint +)

NEGATIVE



A line near letter "C" and **NO LINE** near letter "T" means the patient **DOES NOT** have malaria.



Negative

INVALID RESULT

NO LINE near letter "C" and one or no line near letter "T" means the test is **INVALID**.



Repeat the test using a new RDT if no control line appears.

If no line appears near the letter "C," repeat the test using a **NEW unopened** test packet and a **NEW unopened** lancet.

- 15.** Dispose of the gloves, alcohol swab, desiccant sachet and packaging in a non-sharps waste container.



- 16.** Record the test results in your CHW register. Dispose of cassette in non-sharps waste container



NOTE: Each test can be used **ONLY ONE TIME**. Do not try to use the test more than once.

